



BD EleVation™

Breast Biopsy System

Quick Reference Guide



Device description

The BD EleVation™ Breast Biopsy System is a handheld, self-contained, single insertion, multiple sample biopsy device and is intended to be used with ultrasound guidance.




Loading and locking the BD EleVation™ Probe into the BD EleVation™ Driver


Step 1: Insert probe

The BD EleVation™ Probe may be inserted into the driver in either of the following ways:

OPTION 1: Remove the BD EleVation™ probe from the package using aseptic technique by grasping the BD EleVation™ probe from below. Carefully align BD EleVation™ probe tabs and sample container with corresponding BD EleVation™ Driver slots. Slide the BD EleVation™ probe back to lock into place.




OPTION 2: While the BD EleVation™ Probe remains in the tray packaging, align the BD EleVation™ Driver to the corresponding BD EleVation™ Probe tabs and sample container opening. Slide the BD EleVation™ Driver forward to lock into place.




Step 2: Locking the probe

After loading the probe in one of the following ways shown above, confirm the BD EleVation™ Probe tab is locked into the BD EleVation™ Driver after loading.



Removing the sample

Turn the sample container counterclockwise to remove from the driver.



Remove the tissue samples from the sample container by hinging it open. If additional samples are required, replace the sample container into the device by turning clockwise to lock into place. Ensure the sample container is secure before acquiring additional samples.




Sample container closed



Sample container hinges open


Removing the cannula

The support cannula may be detached and remain in the breast to retain a track to the biopsy site to place a breast tissue marker. To remove the support cannula from the BD EleVation™ probe, depress the coaxial arm and then twist the coaxial hub.



Removing the probe

Remove the BD EleVation™ Probe from the BD EleVation™ Driver by pressing down on the locking tab, sliding the BD EleVation™ Probe cover completely forward and then pulling the BD EleVation™ Probe straight up from the BD EleVation™ Driver.



Cleaning and maintenance

- After each use, wipe the BD EleVation™ Driver with a Super Sani-Cloth or a cloth dampened with water to remove any excess blood or fluids present on the device.

Following the initial wipe, thoroughly clean the BD EleVation™ Driver with Super Sani-Cloths, paying particular attention to the buttons and indicators, locking tab, LED lights, internal gears, and the sides of the driver. Allow the driver to remain wet for a minimum of 3 minutes.

Allow the BD EleVation™ Driver to air dry thoroughly before placing in the wireless charging stand.

- Other than routine cleaning, no other preventative maintenance is required. However, it is recommended to return the device to an authorized Service and Repair Facility once per year for inspection and servicing.
- Do not spray the BD EleVation™ Driver with any fluids. Submerging the BD EleVation™ Driver in fluids may cause it to malfunction and will void the standard warranty.
- Sterilization and exposure to fluids may damage the electrical components of the device. If the BD EleVation™ Driver is inappropriately cleaned, it may cause the BD EleVation™ Driver to malfunction and will void the standard warranty.
- Do not autoclave. Do not heat over 54° (129° F).
- The wireless charging stand may be cleaned with the same methods as the BD EleVation™ Driver.

BD EleVation™ Driver charging

Completely charge the BD EleVation™ Driver prior to first use. The battery must be adequately charged prior to beginning each procedure.

To charge the BD EleVation™ Driver:

1. Ensure the appropriate AC adapter plug is installed on the AC power adapter.
2. Connect the AC adapter cord to the wireless charging stand.
3. Connect the AC adapter plug into a power source. If the wireless charging stand has current, the power light on the front of the wireless charging stand will illuminate green.
4. Place the BD EleVation™ Driver in the wireless charging stand. The BATTERY indicator on the BD EleVation™ Driver will illuminate per the table to indicate the battery is charging.






Time to charge the device from zero to full battery charge






A completely exhausted battery will require up to 12 hours to fully charge.

After every use, and when the driver is not in use, the BD EleVation™ Driver should remain in the charging stand.

Battery status

Battery indicator color	Battery status	Action required
Solid Green 	Battery is sufficiently charged to perform biopsy procedure.	No action needed. BD EleVation™ Breast Biopsy System is ready to perform a biopsy.
Blinking green 	Battery has sufficient charge to complete ONE biopsy procedure.	Proceed with performing the biopsy. After completing the procedure the BD EleVation™ Driver MUST be charged prior to the next biopsy.
Solid Red 	Battery is NOT sufficiently charged to perform a biopsy procedure.	Return BD EleVation™ Driver to charging stand to charge battery.
Battery LED is not illuminated – Driver does not respond when probe is loaded	Battery is completely discharged.	Return BD EleVation™ Driver to charging stand to charge battery.

Troubleshooting

Troubleshooting		
BD EleVation™ Breast Biopsy System	Condition	Action required
<p>BATTERY indicator is solid red</p> 	<p>Low battery</p>	<ul style="list-style-type: none"> Place BD EleVation™ Driver on wireless charging stand to charge battery
<p>ERROR indicator blinks red, PRIME/PIERCE and SAMPLE buttons are solid red</p> 	<p>Error during:</p> <ul style="list-style-type: none"> Calibration PRIME/PIERCE Sampling 	<ul style="list-style-type: none"> Remove BD EleVation™ Probe Place BD EleVation™ Driver on a flat surface with hand away from any moving components Press SAMPLE button
<p>ERROR indicator blinks red, SMART MODE indicator blinks red, PRIME/PIERCE and SAMPLE buttons are solid red</p> 	<p>Smart Mode overloaded</p>	<ul style="list-style-type: none"> Remove BD EleVation™ Probe Place BD EleVation™ Driver on a flat surface with hand away from any moving components Press SAMPLE button
<p>BATTERY indicator blinks red, ERROR indicator blinks red, PRIME/PIERCE and SAMPLE buttons are solid red</p> 	<p>BD EleVation™ Driver placed in wireless charging stand while BD EleVation™ Probe is loaded</p>	<ul style="list-style-type: none"> Remove BD EleVation™ Driver from wireless charging stand Remove BD EleVation™ Probe
<p>BATTERY indicator blinks red, ERROR indicator blinks red, PRIME/PIERCE and SAMPLE buttons are solid red</p> <p>NOTE: Driver is also beeping continuously</p> 	<p>BD EleVation™ Driver placed in wireless charging stand while in error</p>	<ul style="list-style-type: none"> Remove BD EleVation™ Driver from wireless charging stand Remove BD EleVation™ Probe Place BD EleVation™ Driver on a flat surface with hand away from any moving components Press SAMPLE button

BD EleVation™ Breast Biopsy System

Indications for use 1. The BD EleVation™ Breast Biopsy System is indicated to obtain tissue samples from the breast or axillary lymph nodes for diagnostic analysis of breast abnormalities. The BD EleVation™ Breast Biopsy System is intended to provide breast tissue for histologic examination with partial or complete removal of the imaged abnormality. 2. The extent of histologic abnormality cannot be reliably determined from its mammographic appearance. Therefore, the extent of removal of the imaged evidence of an abnormality does not predict the extent of removal of a histologic abnormality, e.g. malignancy. When the sampled abnormality is not histologically benign, it is essential that the tissue margins be examined for completeness of removal using standard surgical procedures.

Contraindications 1. The BD EleVation™ Breast Biopsy System is for diagnostic use only, NOT for therapeutic use. 2. The BD EleVation™ Breast Biopsy System is contraindicated for those patients where, in the physician's judgment, there is an increased risk of complications associated with percutaneous removal of tissue samples.

Warnings 1. Patients who may have a bleeding disorder, or who are receiving anticoagulant therapy, may be at increased risk of complications. 2. As with any biopsy instrument, there is a potential risk for infection. 3. The BD EleVation™ Breast Biopsy System should not be used in a Magnetic Resonance Imaging (MRI) Suite. 4. The BD EleVation™ Breast Biopsy System has not been tested using stereotactic guidance or for use with an MRI. 5. The BD EleVation™ Breast Biopsy System should not be used in an operating room. 6. The BD EleVation™ Breast Biopsy System is not classified as an AP or APG device. 7. The BD EleVation™ Breast Biopsy System is not suitable for use in the presence of flammable anesthetic. 8. The BD EleVation™ Breast Biopsy System is not suitable for use in an oxygen rich environment. 9. The BD EleVation™ Driver must only be used with BD EleVation™ Probes and BD EleVation™ Accessories. 10. All breast biopsies should be performed under ultrasound guidance to confirm the BD EleVation™ Probe's position relative to the target region to be sampled and to help mitigate the occurrence of a false negative biopsy. The BD EleVation™ Breast Biopsy System is intended for use with ultrasound imaging only. 11. The battery may only be replaced or disposed of by an authorized Service and Repair facility. 12. Use only with supplied AC power BD EleVation™ Accessories. Removing the AC adapter plug from wall power shall serve as isolation means. Do not position the AC adapter plug and wireless charging stand such that it is difficult to remove the AC adapter plug from the wall outlet if needed to remove mains power. 13. Do not reuse BD EleVation™ Probe. Reusing the BD EleVation™ Probe bears the risk of cross-patient contamination as biopsy probes, particularly those with long and small lumina, joints, and/or crevices between components, are difficult or impossible to clean once body fluids or tissues with potential pyrogenic or microbial contamination have had contact with the BD EleVation™ Probe for an indeterminable period of time. The residue of biological material can promote the contamination of the BD EleVation™ Probe with pyrogens or microorganisms which may lead to infectious complications. 14. Do not resterilize BD EleVation™ Probe. After resterilization, the sterility of the BD EleVation™ Probe is not guaranteed because of an indeterminable degree of potential pyrogenic or microbial contamination which may lead to infectious complications. Cleaning, reprocessing and/or resterilizing the BD EleVation™ Probe increases the probability that it will malfunction due to potential adverse effects on components that are influenced by thermal and/or mechanical changes.

Precautions 1. The BD EleVation™ Breast Biopsy System should only be used by a physician trained in its indicated use, limitations, and possible complications of percutaneous needle techniques. 2. Do not attempt to remove the cover or modify the device in any way.

Potential Complications 1. Potential complications are those associated with any percutaneous removal/biopsy technique for tissue collection. Potential complications are limited to the region surrounding the biopsy site and include hematoma, lymphedema, hemorrhage, infection, non-healing wound, pain, nerve injury, and tissue adherence to the BD EleVation™ Probe while removing it from the breast. 2. As per routine biopsy procedures, it may be necessary to cut tissue adhering to the BD EleVation™ Probe while removing it from the breast.

Please consult product labels and instructions for use for indications, contraindications, hazards, warnings and precautions and full directions for use.

BD Switzerland Sarl

Terre Bonne Park – A4, Route De Crassier,
17, 1262 Eysins, Vaud, Switzerland
T: +41 21 556 3000

CE
01467

bd.com

© 2021 BD. BD, the BD logo, EleVation and TriConcave are trademarks of Becton, Dickinson and Company or its affiliates. All other trademarks are the property of their respective owners. © 2021 BD. All rights reserved. 0121/6101 BD-26892

